

The following lubricants are materials compatible and washable by use with KARL STORZ fiberscopes.

A longer contact duration as 24 h is not allowable. This means that the fiberscopes should be cleaned thoroughly after use.

Please follow the instructions given by the manufacturer. We recommend the following lubricants free for use regarding the material compatibility:

Supplier	Trade name
Astra GmbH	Xylocain 2%
DeltaSelect GmbH	Gelicain 2%
Dr. Deppe	Instru Gel
Dr. Kade	LidoPosterine Salbe
Farco-Pharma GmbH	Instillagel
Lomapharm	Lidocain Gel 2%
Löns-Apotheke	Endopurin
Pharmazeutische Fabrik Montavit Ges.m.b.H.	Cathejell (with Diphenhydramin)
	Cathejell mit Lidocain

For others as mentioned above KARL STORZ gives no guarantee in case of any damage.